A Radial Time Projection Chamber for α detection in CLAS at JLab

R. Dupré^{a,b,*}, S. Stepanyan^c, M. Hattawy^{a,b}, N. Baltzell^{a,c}, K. Hafidi^a, M. Battaglieri^d, S. Bueltmann^e, A. Celentano^d, R. De Vita^d, A. El Alaoui^{a,f}, L. El Fassi^g, H. Fenker^c, K. Kosheleva^a, S. Kuhn^e, P. Musico^d, S. Minutoli^d, M. Oliver^h, Y. Perrin $^{\rm i},$ B. Torayev $^{\rm e},$ E. Voutier $^{\rm b,i}$

> ^aArgonne National Laboratory, Argonne IL 60439, USA ^b Institut de Physique Nucléaire, CNRS/IN2P3 and Université Paris Sud, Orsay, France ^cJefferson Laboratory, Newport News, VA 230606, USA ^dINFN, Sezione di Genova, 16146 Genova, Italy ^eOld Dominion University, Norfolk, VA 23529, United States ^fUniversidad Técnica Federico Santa María, Casilla 110-V Valparaíso, Chile ^gMississippi State University, Mississippi State, MS 39762-5167 ^hUniversity of Chicago, Chicago, IL 60637, United States ⁱLPSC, Université Grenoble-Alpes, CNRS/IN2P3, Grenoble, France

- Electromagnetic Calorimeters (EC) for identification of electrons, photons and neutrons [5].

For certain experiments the base CLAS system was complimented with ancillary detectors. For example, the

Figure 1: A three dimensional representation of the baseline CLAS setup. The full description is given in the text.

Email address: dupre@ipno.in2p3.fr (R. Dupré)

Abstract

A new Radial Time Projection Chamber (RTPC) was developed at the Jefferson Laboratory to track low-energy nuclear recoils to measure exclusive nuclear reactions, such as coherent deeply virtual Compton scattering and coherent meson production off *He. In 2009, we carried out these measurements using the CEBAF Large Acceptance Spectrometer (CLAS) supplemented by the RTPC positioned directly around a gaseous *He target, allowing a detection threshold as low as 12 MeV for *He. This article discusses the design, principle of operation, calibration methods and performances of this RTPC.

1. Introduction

Until recently, the Thomas Jefferson National Accelerators and the properties of the properties o

^{*}Corresponding author

measurement of the Deeply Virtual Compton Scattering (DVCS) process $(eH \rightarrow e'H'\gamma)$, where H is a nucleon or nuclei) necessitates an upgrade of the photon detection system. Indeed, with a 6 GeV electron beam, the majority of DVCS photons are produced at very forward angles, where the acceptance of the EC was poor. To extend the detection range, an inner calorimeter (IC) was built for the E01-113 experiment in 2005 [6]. The IC was constructed from 424 lead-tungstate (PbWO₄) crystals, covering polar angles between 5° and 15° [7]. To protect the CLAS detector and the IC from the large flux of low energy Møller electrons, a 5 T solenoid magnet was placed around the target to shield the detectors. To detect recoiling α particles from the coherent DVCS on Helium, a new radial time projection chamber (RTPC) was developed to track low energy nuclear fragments. The solenoid field was used to track and measure momentum of particles in the RTPC. The CLAS detector supplemented with both IC and RTPC was used during a three months experimental run [8, 9] in 2009 with a longitudinally polarized electron beam of 130 nA and energy of 6.064 GeV incident on a gaseous ⁴He target.

The original design of the RTPC was developed for the BoNuS experiment at Jefferson Lab which took data with CLAS in 2005 [10]. Significant improvements were made to the RTPC mechanical structure and fabrication technique that both increased the acceptance and reduced the amount of material in the path of the outgoing particles. The enhanced design, used in the 2009 DVCS experiment, is described in section 2 of this paper. The data acquisition system described in section 3 was reconfigured to increase the event readout rate. The calibration methods are discussed in section 4 and the tracking algorithm in section 5. Finally, the overall performance of the RTPC is described in section 6.

2. RTPC design

With a 6 GeV incident electron energy, the recoiling $^4{\rm He}$ nuclei from coherent DVCS have an average momentum around 300 MeV/c (12 MeV kinetic energy). Such low energy α particles are stopped very rapidly, so the RTPC was designed to be as close as possible to the target and fit inside the 230 mm diameter shell and cryostat wall of the solenoid magnet bore.

The new CLAS RTPC is a 250 mm long cylinder of 158 mm diameter, leaving just enough room to fit preamplifiers between the RTPC outer shell and the solenoid. The electric field is directed perpendicularly to the beam direction, such that drifting electrons are pushed away from the beam line. These electrons are amplified by three layers of cylindrical gas electron multipliers (GEM) and detected by the readout system on the external shell of the detector as illustrated in the Fig. 2. The RTPC is segmented into two halves with independent GEM amplification systems that cover about 80% of the azimuthal angle.




Figure 2: Schematic drawing of the CLAS RTPC in a plane perpendicular to the beam direction. See text for description of the elements.

We detail here the different regions shown in Fig. 2 starting from the beam line towards larger radius:

- The 6-atm 4 He target extends along the beam axis; it is a 284-mm-long, 6-mm-diameter Kapton straw with a 27- μ m wall.
- The first gas gap covers the radial range from 3 mm to 20 mm. It is filled with $^4{\rm He}$ gas at 1 atm to minimize secondary interactions from Møller electrons scattered by the beam. This region is surrounded by a 4 μ m thick window made of grounded aluminized Mylar.
- The second gap region extends between 20 mm and 30 mm and is filled with the gas mixture of 80% neon (Ne) and 20% dimethyl ether (DME). This region is surrounded by a 4 μ m thick window made of aluminized Mylar set at -4260 V to serve as the cathode.
- The drift region is filled with the same Ne-DME gas mixture and extends from the cathode to the first GEM, 60 mm away from the beam axis. The average electric field in this region is perpendicular to the beam and about 550 V/cm.
- The electron amplification system is composed of three GEMs located at radii of 60, 63 and 66 mm. In this configuration, the first GEM layer serves as the anode and each subsequent GEM is set at a lower voltage to obtain a strong (~1600 V/cm) electric field between the GEM foils. A 275 V bias is applied across each GEM for amplification.
- The readout board has an internal radius of 69 mm and collects charge from the GEMs. Pre-amplifiers




Figure 3: Image of a typical GEM foil similar to the one used for our RTPC [11].

are plugged directly on its outer side and transmit the signal to the data acquisition electronics.

The GEM technology has been chosen for the flexibility of the GEM foils, which can be easily used to produce a curved amplification surface. Also, GEMs are known to have relatively low spark rate [12], which is important when trying to detect highly ionizing slow nuclei that deposit large amount of energy. The GEMs for this RTPC are made from a Kapton insulator layer, 300 μ m thick, sandwiched between two 5 μ m copper layers¹. The mesh of each GEM layer is chemically etched with 50 μ m diameter holes with double-conical shapes as illustrated in Fig. 3. The potential difference applied between the two copper layers of the GEM creates a very strong electric field in each hole leading to high ionization and amplification.

The drift gas used in the experiment is a 80-20% Ne-DME mixture. This choice has been made in order to balance the energy deposit, which is critical for proper particle identification, with a reasonable Lorentz angle. Calculations using the MAGBOLTZ program [13] showed that with the axial 5 T field, we would have a Lorentz angle of about 23° with this gas mixture.

One of the important challenge in developing the radial TPC was to obtain a good support structure for the GEM foils to allow a tractable installation of the GEMs. At the same time, we wanted to keep the material budget small in the forward region where we detect other particles in subsequent detectors. We successfully realized these prerequisites by using fiber glass rings glued to each end of the GEM foils to form self supporting cylinders that could be installed independently in the RTPC after gluing and soldering operations. The rigidity of the GEM foils was enough for the structure to be self-supporting and only the upstream end of the cylinder was fixed to the main mechanical support structure. This design only left a light fiberglass ring in the downstream end, reducing to a minimum secondary interactions.




Figure 4: A schematic representation of the readout pads. The shaded sixteen pads are a group of pads that are connected to the same pre-amplifier.

3. Readout System

The RTPC electron collection system had 3200 readout pads. These elements were located at the end of the amplification region, 69 mm from the central axis. Figure 4 illustrates the configuration of the 5 by 4.45 mm pads, where the shift between the rows was implemented to reduce aliasing. Each half of the RTPC had 40 rows and 40 columns of pads. The shaded region in Fig. 4 shows how pads were grouped to 16 channels pre-amplifier boards. The pre-amplifier boards, already employed in the BoNuS RTPC [10], serve the dual purpose of inverting the RTPC signals polarity – from negative to positive – to match the requirements of the subsequent readout system, and driving the 6-m long ribbon cable that connects to it.

The readout system, similar to the original BoNuS RTPC system [10], was based on the front end electronic boards originally developed for the ALICE TPC readout system at CERN [14]. The Front End Card (FEC) hosted the analogue receiver circuit and the subsequent amplification and digitization stage, based on the PASA [15] and ALTRO [16] ASICS. A Readout Control Unit (RCU) board was used to distribute the trigger signal to FECs and for data readout, with each RCU handling up to 25 FECs. Communication between readout controller application (ROC) and FECs was performed trough a custom backplane, implementing a low-voltage signal bus. The RCU communicated with the CLAS DAQ system trough a 200 MB/s optical link, connected to a data acquisition PC hosting a ReadOut Receiver Card (RORC). This PC hosted the ROC used to interface with the CLAS DAQ system. An Ethernet link was also present, for slow-controls and monitoring.

¹The GEM foils were produced by Tech-Etch, Inc.




Figure 5: Schematic representation of the RTPC readout system, showing the different elements of the front end electronics.

The standard CLAS electron trigger was used to initiate the RTPC readout. For each event, 100 samples/channel were digitized and processed by the ALTRO ASIC. In order to reduce the data size, ALTRO was operated in zero-suppression mode, keeping only samples from channels above a programmable threshold set just above the noise level. Samples were pre-processed by a glitch-filter to reject noise and spurious pulses. To properly reconstruct the signal shape, $N_{PRE} = 3$ samples before threshold-crossing and $N_{POST} = 3$ samples after the signal returns below threshold were saved.

The trigger signal also initiates RCU readout operation from FEC boards. All the measured samples from active channels are reported, together with a channel identifier and a timestamp, to the ROC application, that in turns sends them on the main event builder. Readout from FECs occurs in "block-transfer" mode, making use of the AL-TRO internal multi-event buffer. This feature could be exploited thanks to the new RCU boards developed for the RTPC readout system². In order to read all the detector readout pads, four FEC crates were used, each equipped with 6 boards, plus a ROC. A schematic of the readout system, for a single crate, is reported in figure 5. This configuration permitted to reduce the dead time associated to FEC readout operations, which is scaling linearly with the number of boards in the crate. During the 2009 run, the system was successfully operated with a DAQ rate of 3.1 kHz and a live time of 70%, for a luminosity of about 10^{34} cm⁻².s⁻¹ and a beam energy of 6.067 GeV.

Finally, during data reconstruction, the acquired samples were processed to obtain, for each readout pad, the

accumulated charge (ADC) and the pulse time (T). Since pulse time was obtained as the time-stamp of the first sample above threshold, referred to the trigger time, the resolution is equivalent to the ALTRO sampling time, 100 ns

4. Calibration

The timing information collected for each signal above threshold was used to infer the origin of the ionization electrons and then the trajectory of the detected particle. The recorded ADCs were used to reconstruct the deposited energy per unit of length $\left(\frac{dE}{dX}\right)$ which, together with the momentum calculated from the trajectory, enabled the particle identification.

In this section we will detail the methods used to calibrate the drift speed, drift paths and gains of the detector. Drift speed and paths were initially calculated using the MAGBOLTZ [13] program, then refined using data to account for variations of the run conditions. We always assume cylindrical symmetry in the chamber for the calibration, such that none of the parameters depend on the azimuthal angle ϕ . The initial MAGBOLTZ calibration was improved through several iterations of the process described below, with each time an increasing number of tracks reconstructed in the RTPC. The figures presented in this section are the ones obtained while performing the last iteration of this calibration process.

4.1. Drift Speed Parametrization

After the ionization, the released electrons drift to the cylindrical detection plane under the effect of the electric field. The electrons released close to the cathode take the most time to reach the readout pads, the cylindrical symmetry insures that this maximum drift distance is the

²In the original BONUS system, the U2F readout controller had no multi-event capability and used a USB connection rather than fiber optics for readout. These factors limited the maximum readout rate from the BONUS detector to about 500 Hz.




Figure 6: A schematic drawing of a ⁴He track (in green) traversing the drift region, with the drift paths followed by the electrons (in black).




Figure 7: Time distribution of the hits associated with a track in one experimental run. Trigger time is defined as T=15, the time unit is the ALTRO sampling time of 100 ns.

same for all tracks. We illustrate this in Fig. 6, where we depict a typical ${}^{4}\text{He}$ track. By measuring the maximum time (T_{Max}) , we can therefore infer the drift speed of the electrons in the RTPC.

To measure the maximum drift time, we used the time profile of hits from identified tracks shown in Fig. 7. We can clearly observe the dropping edge expected from geometrical considerations. We define a value $T_{Max/2}$ at which the dropping edge passes half the maximum number of hits in the histogram. This value was measured in bins along the 200 mm RTPC's length to take into account variations in the electric and magnetic field in the RTPC (see Fig. 8).

Due to the non perfect experimental conditions, in particular possible contamination of our gas mixture, the drift speed changed during the three months long experimental




Figure 8: Maximum time of collected hits as a function of the track position on the z-axis for one experimental run.

run³. Figure 9 shows the $T_{Max/2}$ values for individual runs (approximately 2 hours long). We observe significant change in the drift speed before and after run 61600, while variations within these periods are 2%.

In summary, we obtain from our calibration a parametrization of the drift speed as a function of both position along the beam axis and the run number. These functions were extracted for our entire data set and implemented in the track reconstructions code.

4.2. Drift Paths Calibration

The drift path is the trajectory followed by the electrons released through ionization in the gas. We calculated them with MAGBOLTZ [13], but it requires knowledge of the detector's geometry, gas mixture composition, and of course the electric and magnetic fields over the whole volume of the detector. We used this calculation as a first calibration, but, as can be seen with the drift speed, conditions in the chamber were changing over time. Moreover, the 4 μ m foil used as a cathode is easily deformed, such that we expect the geometrical accuracy to be of few millimeters, directly impacting our knowledge of the electric field. These problems, already encountered for the BoNuS RTPC calibration [10], motivated the acquisition of specific calibration runs. These were taken with a lower energy electron beam (1.204 and 1.269 GeV) to enhance the cross section of the elastic scattering ($e^4\text{He}\rightarrow e^4\text{He}$). In this process, the measurement of the electron kinematics allows to calculate directly the kinematics of the Helium nucleus. By comparing the calculated momentum and angle of the recoil alpha particle to the measurement in the RTPC, we tuned the drift paths independently of our knowledge of the chamber's conditions.

³We had to increase the gas flow during the experiment due to a small leak in the RTPC, which concurred to the large shift of drift time observed around run number 61600. While our gas system was kept slightly over atmospheric pressure to limit contamination from air or other external gases, it is likely that this leak was the source of modification of the drift speed.




Figure 9: $T_{max/2}$ versus the experimental run numbers.




Figure 10: $\Delta\phi$ versus T distribution for tracks from one bin in longitudinal position along the RTPC. The black line represents the final drift paths in this bin.

Based on the kinematics of the electrons in the calibration data, we generated the Helium nucleus in our RTPC GEANT4 simulation [17]. Then, we compared the calculated GEANT4 trajectory of the Helium nuclei to the hits measured in the chamber. To perform this drift path extraction, we made a first approximation assuming a linear dependence between the radius of emission of the charge and its time of detection, and then refined our result. Indeed, because of the magnetic field, the drift paths are not linear in the RTPC, but the curvature was minimal and this process converged already on the second iteration.

At the end of the extraction procedure, the azimuthal difference between the detection pad and the ionization point $(\Delta\phi)$ was extracted as a function of time. In Fig. 10, we show the resulting data points for one bin, where the drift path is easily identified and eventually fitted for implementation in our reconstruction codes.

To verify the stability of the drift paths, this procedure was carried out using both the 1.204 GeV data from the beginning of the run period and the 1.269 GeV data from the end of the run period (shown in blue on Fig. 9). Interestingly, we found very similar drift paths for the two data sets and concluded that any changes in the system only significantly affected the drift speed.

4.3. Gain Calibration


To calibrate the gains, we compared the experimental ADCs to the energy deposited for each pad individually in GEANT4 by similar simulated tracks (using the same elastic events as for the drift paths calibration). This requires a very good GEANT4 simulation including drift paths, but also the spread of the charge along the path before reaching the pad, so that the simulated hits match the experimental ones. Moreover, the simulation has to match the DAQ features that can lead to cutting out hits. After setting the simulation properly, we compared simulation to experiment on an event by event basis as shown in Fig. 11. The gain for each pad was calculated as the ratio of the measured ADCs to the simulated deposited energy. Then, these gains were refined using correction factors obtained from a sample of good tracks. For each track, we calculate the corrected energy deposit on a pad and compared it to the average deposit recorded by the other pads. This last step was implemented to avoid any bias from this simulation based calibration method. Final results are shown in Fig. 12, where energy loss of particles is plotted against momentum over charge ratio. One can clearly see the band for ⁴He in its expected position.

5. Track Reconstruction

5.1. Noise Rejection

Two independent noise signatures were identified in the raw data and removed in software prior to track reconstruction. Both are transient and isolated to a subset of the readout channels.

The first is an oscillatory noise located early in the readout time window, shown in the top panel of Fig. 13 for a particularly noisy channel. Its amplitude is similar to those of real tracks. About 18% of the readout channels exhibit large contributions from this noise characteristic. Due to its unique time-energy correlation for the given channels, the noise could be removed on an event by event and channel by channel basis without significant loss of






Figure 11: Simulated (upper) and experimental (lower) ADC and T distributions of a track. The colors indicate the pads, same color in top and bottom indicate that they are the same pad.

good signals. The result of the procedure is illustrated in the bottom panel of Fig. 13.

The second noise signature was a coherent noise affecting about 25% of the pre-amplifiers boards, when simultaneous hits in most of the 16 channels of the board were recorded. An event-based technique to identify and remove this noise was developed based on counting simultaneous hits in each pre-amplifier group, and, if sufficiently large, perform a dynamic pedestal subtraction based on the average ADC of neighboring channels within this group.

The sources of these effects were not determined, but rejection techniques allowed to reconstruct 10% more good tracks and recover 70 channels that were previously ignored due to excessive noise levels.

5.2. Track Fitting

The tracking starts with reconstructing the spacial origin of the hits using the extracted drift speed and drift path parameters. For each registered hit, we calculate a position of emission from the signal time and the pad position. The third step is to create chains of hits. The maximum distance between two close adjacent hits has to be less than 10.5 mm to chain them, which roughly corresponds to neighbors and next to neighbors. We fit the chains with a helix if they have a minimum of 10 hits. We then eliminate from the chain the hits that are 5 mm or farther from the fit, as they are not likely part of the same

track. This new reduced chain is used for a second and final helix fit.

For energy deposition, the mean $\frac{dE}{dx}$ is calculated as

$$\left\langle \frac{dE}{dX} \right\rangle = \frac{\sum_{i} \frac{ADC_{i}}{G_{i}}}{L},\tag{1}$$

where the sum runs over all the hits of the track, G_i is the gain of the associated pad, and L is the visible track length in the active drift volume.

5.3. Energy Loss Corrections

Energy loss between the target and drift region was significant in the RTPC and necessitated a correction for optimal momentum reconstruction at the primary interaction vertex. The dominant loss was in the 27- μ m-thick Kapton target wall, with significant contributions also from the pressurized target gas and the foils before reaching the drift region. Corrections were developed based on GEANT4 simulations with the full RTPC geometry and parameterized in terms of recoil curvature in the drift region and polar angle, separetely for all recoil hypotheses (p, d, 3 H, 3 He, 4 He). At our average coherent 4 He DVCS kinematics, energy losses were about 5 MeV, while for e^{-4} He elastic scattering losses were about 3 MeV, which corresponds to momentum corrections of 25% and 15%, respectively.

6. Performance Studies

The primary data sample used for calibration and performance assessment of the RTPC was elastic scattering with a 1.2 GeV electron beam. The electron's momentum and direction is measured with CLAS, which uniquely determines the expected recoiling 4 He kinematics. Matching requirements between reconstructed and expected z-vertex and direction of the RTPC track provides a clean selection of elastically-scattered 4 He, shown in Fig. 14.

6.1. Resolution

Elastic scattering was used to estimate the tracking resolution of the RTPC based on the residual between the expected and measured $^4\mathrm{He}$ tracks. The RTPC resolutions, after removing contributions from the electron, are shown in Table 1, and are very similar for the two halves of the RTPC. Note that the θ - and z-resolutions are highly correlated.

	σ_z	$\sigma_{ heta}$	σ_{ϕ}	σ_p
Left	5.3 mm	3.8°	1.9°	9%
Right	$6.5~\mathrm{mm}$	4.0°	1.9°	8%

Table 1: The resolutions of the two modules of the RTPC for z-vertex, polar and azimuthal angles, and momentum.




Figure 12: $\frac{dE}{dX}$ vs. p/q distributions for the left (on the left) and for the right (on the right) half of the RTPC after gain calibration. The lines are theoretical expectations from the Bethe-Bloch formula for 4 He (red), 3 He (black), 3 H (blue), 2 H (pink) and protons (gray).




Figure 13: The ADC vs. T spectrum for an example noisy channel before (top) and after (bottom) noise rejection algorithms. Only hits associated with tracks are included, and the selection of events and tracks is the same in both plots.

6.2. Efficiency

We measured the efficiency of the RTPC using elastic scattering on ⁴He by comparing the inclusive yield, based only on electron detection, to the exclusive elastic yield, where the Helium recoil is also detected (see Fig. 14). We present in Fig. 15 the results for the two halves of the detector. We observe that the left and the right modules have similar efficiencies except near the upstream target window. This difference is due to the large number of




Figure 14: The W distribution calculated from electron kinematics before ("inclusive") and after ("exclusive") requiring a matching track in the RTPC.

dead channels concentrated in this part of the left half of the detector

7. Conclusion

We reported on the construction, operation and calibration of a small RTPC designed to measure Helium-4 nuclei in high rate environment. The operation of the detector was successful and allowed to detect Helium nuclei with a 75% efficiency and a readout rate of 3.1 kHz triggered by the detection of high energy electrons and photons in the CLAS spectrometer.

8. Acknowledgments

The authors thank the staff of the Accelerator and Physics Divisions at the Thomas Jefferson National Accelerator Facility who made this work possible. This work




Figure 15: The RTPC 4 He detection efficiency as a function of the longitudinal position along the detector.

was supported in part by the French Centre National de la Recherche Scientifique (CNRS), the Italian Istituto Nazionale di Fisica Nucleare (INFN) and the U.S. Department of Energy. The author (M. Hattawy) also acknowledges the support of the Consulat Général de France à Jérusalem. The Southeastern Universities Research Association operates the Thomas Jefferson National Accelerator Facility for the United States Department of Energy under contract DE-AC05-06OR23177.

References

- B.A. Mecking et al., The CEBAF large acceptance spectrometer, Nucl. Inst. and Meth. A 503, 513 (2003).
- [2] M.D. Mestayer et al., The CLAS drift chamber System, Nucl. Inst. and Meth. A 449, 81 (2000).
- [3] G. Adams et al., The CLAS Cerenkov detector, Nucl. Inst. and Meth. A 465, 414 (2001).
- [4] E.S. Smith et al., The time-of-flight system for CLAS, Nucl. Inst. and Meth. A 432, 265 (1999).
- [5] M. Amarian et al., The CLAS forward electromagnetic calorimeter, Nucl. Inst. and Meth. A 460, 239 (2001).
- [6] F.X. Girod et al. (CLAS Collaboration), Measurement of Deeply virtual Compton scattering beam-spin asymmetries, Phys.Rev.Lett. 100 (2008) 162002
- [7] Hyon-Suk Jo, Etude de la Diffusion Compton Profondément Virtuelle Sur le Nucléon avec le Détecteur CLAS de Jefferson Lab: Mesure des Sections Efficaces polarisées et non polarisées, IPNO-Thesis, 2007.
- [8] G Asryan et al., Meson spectroscopy in the Coherent Production on ⁴He with CLAS, Jlab proposal to PAC31 (2007).
- [9] K. Hafidi et al., Deeply virtual Comton scattering off ⁴He, Jlab proposal to PAC33 (2008).
- [10] Howard C. Fenker et al., BoNuS: Development and Use of a Radial TPC using Cylindrical GEMs, Nucl. Instrum. Meth. A592, 273-286 (2008)
- [11] F. Sauli, The Gas Electron Multiplier (GEM): Operating Principles and Applications, Nucl. Instr. and Meth, A805(2016)2
- [12] J. Beringer et al. (Particle Data Group), Particle detectors at accelerators, Phys. Rev. D 86, 010001, pages: 339-368, 2012.
- [13] S. Biagi, Monte Carlo simulation of electron drift and diffusion in counting gases under the influence of electric and magnetic fields, Nucl. Inst. and Meth. in Phy. Res. A, vol. 421, pp. 234-240, 1999
- [14] L. Musa et al., The ALICE TPC front end electronics, 2003 IEEE Nuclear Science Symposium, 2003, pp. 3647-3651 Vol.5

- [15] H. K. Soltveit et al., "The Preamplifier shaper for the ALICE TPC-Detector," Nucl. Instrum. Meth. A 676 (2012) 106
- [16] , R. E. Bosch, et al., "The ALTRO chip: a 16-channel A/D converter and digital processor for gas detectors", IEEE Transactions on Nuclear Science, vol. 50, no. 6, pp. 2460-2469, Dec. 2003
- [17] http://geant4.cern.ch